

BZ-R6

INFRARED THERMOMETER

SUNPHOR™



## FEATURES

- Contactless Infrared Thermometer
- Scan Distance 5-15 cm
- Instant Temperature Reading (0.5 Seconds)
- LCD Screen with Backlight Display
- 3 Color for Different Temperature Reading
- Audible & Visual Alarm
- Suitable for Body, Water, Room, and Milk Temperature

## SPECIFICATION

- Measuring Range : 32.0C (90F)~ 43.0C (109.4F)
- Display Resolution : 0.1 C / 0.1 F
- Measuring Distance : 5 to 15 cm (2 to 5.9in)
- Temperature Unit : C or F
- Operating Temperature : 5 C (41 F) ~ 40 C ( 104 F)  
RH:<85%
- Storage Temperature : -20 C (-4 F) ~ 55 C (131 F)  
RH:10% ~ 85 %
- Power : DC 3V (2 x AAA Battery)
- Dimension : 160mm x 90mm x 45mm (L\*W\*H)
- Package Dimension : 197.5mm x 122.5mm x 50.0mm (L\*W\*H)
- Product Weight : 100g (Batteries not included)
- Automatic Shutdown : 10 Seconds

**AUTHORIZED DEALER :**

## CERTIFICATE

of Conformity

EC Council Directive 2014/30/EU

Certificate No.: SiCT200408130C

- Applicant** : Henan Bingzun Industrial Co., Ltd  
50 meters next to the south of Guihua Villa, Xin'an Industry Cluster District, Luoyang City, Henan Province, China
- Manufacturer** : Henan Bingzun Industrial Co., Ltd  
50 meters next to the south of Guihua Villa, Xin'an Industry Cluster District, Luoyang City, Henan Province, China
- Product** : Non-contact Electronic Thermometer
- Trademark** : SUNPHOR
- Model** : BZ-R6  
BZ-R3,R5,R6,R7,R8,R9,R11
- Test Standard** : EN 55014-1:2017  
EN 61000-3-2:2019, EN 61000-3-3:2013+A1:2019  
EN 55014-2:2015

The certificate of conformity is based on an evaluation of a sample of the above mentioned product. Technical report is at the applicant disposal. This is to certify the tested sample that is in conformity with all provisions of above EMC directive. It is only valid in connection with the test report number SiCT200408130E.


The certificate does not imply the assessment of the production and does not permit using the SiCT's logo without permission.



Authorized Signer: \_\_\_\_\_

Jesse Liu/Manager

Date: Apr. 08, 2020

**Shenzhen SiCT Technology Co., Ltd.**

4F, Building A, EDerH Industrial District, Fukang Community, Longhua

Subdistrict, Longhua District, Shenzhen, Guangdong, P. R. China

E-mail: info@sict-lab.com Http:// www.sict-lab.com.cn

## Supplier's Declaration of Conformity

Certificate No.: SiCT200408132C

- Applicant** : Henan Bingzun Industrial Co., Ltd  
50 meters next to the south of Guihua Villa, Xin'an Industry Cluster  
District, Luoyang City, Henan Province, China
- Manufacturer** : Henan Bingzun Industrial Co., Ltd  
50 meters next to the south of Guihua Villa, Xin'an Industry Cluster  
District, Luoyang City, Henan Province, China
- Product** : Non-contact Electronic Thermometer
- Trademark** : SUNPHOR
- Model** : BZ-R6  
BZ-R3,R5,R6,R7,R8,R9,R11
- Test Standard** : FCC Part 15, Subpart B  
ANSI C63.4: 2014

The certificate of conformity is based on an evaluation of a sample of the above mentioned product. Technical report is at the applicant disposal. This is to certify the tested sample that is in conformity with all provisions of above FCC standard. It is only valid in connection with the test report number SiCT200408132E.

The certificate does not imply the assessment of the production and does not permit using the SiCT's logo without permission.



Authorized Signer: \_\_\_\_\_

Jesse Liu/Manager

Date: Apr. 03, 2020



**Shenzhen SiCT Technology Co., Ltd.**

4F, Building A, EDerH Industrial District, Fukang Community, Longhua  
Subdistrict, Longhua District, Shenzhen, Guangdong, P. R. China

# NOTICE OF COMPLETION

To:

**Henan Bingzun Industrial Co., Ltd**  
**50 Meters Next To The South Of Guihua Villa, Xin'An Industry Cluster District**  
**Luoyang, Henan, 471000, CHINA**

is registered and has listed the following medical device with **the U.S. Food and Drug Administration(FDA)**:

**Owner/Operator Number** : 10073151

**Listing Number**: D400073

**Product Code** : FQZ

**Product** : Non-contact Infrared Thermometer


**Model(s)** : BZ-R3, R5, R6, R7, R8, R9, R11

**Date Of Registration Status**: 2020

Registrar Corp will confirm that such registration remains effective upon request and presentation of this certificate until the end of the year stated above, unless said registration is terminated after issuance of this certificate. Registrar Corp makes no other representations or warranties, nor does this certificate make any representations or warranties to any person or entity other than the named certificate holder, for whose sole benefit it is issued. This certificate does not denote endorsement or approval of the certificate-holder's device or establishment by **the U.S. Food and Drug Administration(FDA)**. Registrar Corp assumes no liability to any person or entity in connection with the foregoing.

Pursuant to 21 CFR 807.39, "Registration of a device establishment or assignment of a registration number does not in any way denote approval of the establishment or its products. Any representation that creates an impression of official approval because of registration or possession of a registration number is misleading and constitutes misbranding."

**The U.S. Food and Drug Administration(FDA)** does not issue a certificate of registration, nor does the U.S. Food and Drug Administration recognize a certificate of registration. Registrar Corp is not affiliated with **the U.S. Food and Drug Administration(FDA)**



Chief Engineer / Zaccheo Lee

Issued: May 07, 2020

Validity Period:2020-12-31

Establishment Registration & Device Listing Website :

<https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfRL/rl.cfm>

# CERTIFICATE

## of Conformity

**EC Council Directive 2011/65/EU and 2015/863/EU**

**Certificate No.: SiCT200408131C**

- Applicant** : Henan Bingzun Industrial Co., Ltd  
50 meters next to the south of Guihua Villa, Xin'an Industry Cluster  
District, Luoyang City, Henan Province, China
- Manufacturer** : Henan Bingzun Industrial Co., Ltd  
50 meters next to the south of Guihua Villa, Xin'an Industry Cluster  
District, Luoyang City, Henan Province, China
- Product** : Non-contact Electronic Thermometer
- Trademark** : SUNPHOR
- Model** : BZ-R6  
BZ-R3,R5,R6,R7,R8,R9,R11
- Test Standard** : IEC62321-1:2013, IEC62321-3-1:2013, IEC62321-4:2013  
IEC62321-5:2013, IEC 62321-6:2015, IEC 62321-7-1:2015  
IEC 62321-7-2:2017, IEC 62321-8:2017

The certificate of conformity is based on an evaluation of a sample of the above mentioned product. Technical report is at the applicant disposal. This is to certify the tested sample that is in conformity with all provisions of above RoHS 2.0 Directive (EU) 2015/863 and (EU)2017/2102 amending Annex II to Directive 2011/65/EU . It is only valid in connection with the test report number SiCT200408131R.

The certificate does not imply the assessment of the production and does not permit using the SiCT's logo without permission.

# RoHS

Authorized Signer: \_\_\_\_\_

Jesse Liu/Manager

Date: Apr. 08, 2020



**Shenzhen SiCT Technology Co., Ltd.**

4F, Building A, EDerH Industrial District, Fukang Community, Longhua

Subdistrict, Longhua District, Shenzhen, Guangdong, P. R. China

E-mail: info@sict-lab.com Http:// www.sict-lab.com.cn